

Program Notice

FGIS-PN-19-09

June 10, 2019

**SAMPLE COLLECTION RESPONSIBILITIES FOR VERIFYING
THE ACCURACY OF MOISTURE METER CALIBRATIONS CROP YEAR 2019**

Contents

1. PURPOSE2

2. BACKGROUND.....2

3. EFFECTIVE DATE2

4. REPLACEMENT HIGHLIGHTS.....2

5. RESPONSIBILITIES.....2

6. ASSIGNMENTS FOR SELECTED FIELD OFFICES2

7. INSTRUCTIONS.....3

The U.S. Department of Agriculture (USDA) prohibits discrimination in its programs on the basis of race, color, national origin, sex, religion, age, disability, political beliefs, and marital or familial status. (Not all prohibited bases apply to all programs.) Persons with disabilities who require alternate means for communication of program information (Braille, large print, audiotape, etc.) should contact USDA’s TARGET Center at (202) 720-2600 (voice and TDD).

To file a complaint, write to the USDA, Office of Civil Rights, Room 326-W, 1400 Independence Avenue, SW, Washington, DC 20250-9410, or call (202) 720-5964 (voice and TDD). USDA is an equal employment opportunity employer.

1. PURPOSE

This program notice transmits collection assignments for samples needed to verify the accuracy of official moisture meter calibrations. It also restates the procedure for collecting and submitting samples.

2. BACKGROUND

The annual Moisture Meter Calibration Study is conducted on current year crop samples to assess the accuracy of the official inspection system and of National Type Evaluation Program (NTEP)-certified moisture meters. Each year, the evaluation is performed on samples submitted to the Inspection Instrumentation Branch (IIB) from the field offices and official service providers. Sample collection assignments for the respective offices are based on three years of crop production data within the geographic areas of responsibility.

Calibrations will be verified over the working moisture ranges but there is a significant need for samples that extend the moisture ranges shown in Directive 9180.61, Official Moisture Calibrations for UGMA-Compatible Meters. While it is understood that all requested moisture levels may not be available in all areas every year, field offices should make all reasonable efforts to provide the requested number of samples in each moisture range and to find samples at the moisture extremes requested.

3. EFFECTIVE DATE

This program notice is effective upon receipt for the 2019 crop production. Wheat samples should be submitted by October 19, 2019, and all other grain samples by December 21, 2019.

4. REPLACEMENT HIGHLIGHTS

This program notice supersedes FGIS PN-18-06, dated May 3, 2018.

5. RESPONSIBILITIES

The collection and submission of samples for the annual Moisture Meter Calibration Study are considered regular duties of the selected field offices. All associated time will be charged to the field office standardization management code.

6. ASSIGNMENTS FOR SELECTED FIELD OFFICES

The 2019 sample request is similar to the 2018 request which was reduced for some grains from previous years, due to the improved capability of the UGMA moisture meters. With this reduction, it becomes significantly more important that each office give their best effort to fulfill their request as stated in Table 1, so that the calibrations can be maintained with the same confidence as in the past.

During the 2019 growing season, the indicated numbers of samples of the commodities listed in Table 1 must be collected, tested for moisture, and submitted by the respective field offices to TSD-IIB. Each sample should weigh approximately 1500 grams.

7. INSTRUCTIONS

- a. The purpose of this effort is to obtain representative samples from the entire nation. Therefore, it is important to have each office fill its quota at all moisture levels, if possible. However, do not submit extra samples in any moisture range, and do not adjust the moisture level of samples by adding water or by drying in the laboratory.
- b. Samples with moisture levels beyond the established moisture ranges are valuable for extending these ranges. For this reason, some of the ranges of requested samples (Table 1) have been extended beyond established limits. When submitting samples, if the moisture falls outside the range of the applicable official moisture meter calibration, obtain an approximate moisture. The true moisture will later be determined at TSD by air oven.
- c. If dockage is removed for inspection purposes, do not recombine it before submitting the sample.
- d. The significant amount of time and effort invested in collecting and submitting the moisture samples can easily be lost through insect damage, microbial spoilage, or late sample submission. To prevent such loss, please collect the samples during the growing season and at harvest time and submit them promptly. Samples above 16 percent (above 14 percent for sunflower seeds and 11 percent for minor oilseeds) require special handling. To minimize loss by spoilage, keep high moisture samples refrigerated (not frozen) until shipped and ship the samples by UPS at least 48 hours before a weekend/holiday.
- e. An easy way to account for samples submitted is to prepare mailing tags [Attachment 1] for the total number of samples of each commodity to be collected. On the back of each tag, write the commodity and moisture range. When all of the mailing tags are used, the collection assignment has been met.
- f. The UGMA-Compatible moisture meters have a built-in test weight correction. These corrections need to be checked using external test weight data. For samples of sufficient volume, test weight will be determined by TSD-IIB, so it is not necessary to record test weight on the mailing tag. However, some submitted samples are too small to fill the kettle. For such samples, please record the test weight on the tag (or transmittal slip) if it is known.
- g. Questions concerning these instructions should be directed to (816) 891-0489. If there is a special problem with a sample assignment, please notify the Moisture Laboratory at (816) 891-0445, as early in the season as possible.

- h. Seal each sample in a polyethylene bag (6 mil thickness) and insert the bag into a canvas grain bag. When shipping several samples in a larger container (box or mail sack), a canvas grain bag around each poly bag will help prevent the poly bags from breaking in transit. Record the field office location, date, commodity, official meter moisture, and test weight (if sample size is limited) on the back of the mailing tag or transmittal form [Attachment 1] accompanying the sample. Attach the mailing tag to the bag. Send samples to:

USDA AMS-NGC
Technology & Science Division
Moisture Laboratory
10383 N. Ambassador Drive
Kansas City, MO 64153-1394

Attachments

Attachment 1: Moisture Sample Transmittal Form

Moisture Sample Transmittal Form

Field Office Use Only:

OFFICE _____ MOISTURE _____

DATE _____ TEST WT. _____

COMMODITY _____

TSD Use Only: *Date Received*

Moisture Sample Transmittal Form

Field Office Use Only:

OFFICE _____ MOISTURE _____

DATE _____ TEST WT. _____

COMMODITY _____

TSD Use Only: *Date Received*

Moisture Sample Transmittal Form

Field Office Use Only:

OFFICE _____ MOISTURE _____

DATE _____ TEST WT. _____

COMMODITY _____

TSD Use Only: *Date Received*

Moisture Sample Transmittal Form

Field Office Use Only:

OFFICE _____ MOISTURE _____

DATE _____ TEST WT. _____

COMMODITY _____

TSD Use Only: *Date Received*

Moisture Sample Transmittal Form

Field Office Use Only:

OFFICE _____ MOISTURE _____

DATE _____ TEST WT. _____

COMMODITY _____

TSD Use Only: *Date Received*

Moisture Sample Transmittal Form

Field Office Use Only:

OFFICE _____ MOISTURE _____

DATE _____ TEST WT. _____

COMMODITY _____

TSD Use Only: *Date Received*

Moisture Sample Transmittal Form

Field Office Use Only:

OFFICE _____ MOISTURE _____

DATE _____ TEST WT. _____

COMMODITY _____

TSD Use Only: *Date Received*

Moisture Sample Transmittal Form

Field Office Use Only:

OFFICE _____ MOISTURE _____

DATE _____ TEST WT. _____

COMMODITY _____

TSD Use Only: *Date Received*

Table 1. Sample collection assignments, 2019 Crop Year

1. Barley, Six-Rowed	Office	<u>7-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-21</u>	<u>All</u>		
	DIOO	14	14	14	14	4	60		
2. Barley, Two-Rowed	Office	<u>7-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-21</u>	<u>All</u>		
	Washington	2	3	3	2	2	12		
	DIOO	8	12	12	12	4	48		
3. Corn	Office	<u>6-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-18</u>	<u>18-22</u>	<u>22-26</u>	<u>26-32</u>	<u>All</u>
	Toledo	2	5	6	5	5	4	3	30
	DIOO	11	26	29	28	27	26	13	160
4. Oats	Office	<u>8-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>All</u>			
	DIOO	5	5	3	2	15			
	Special Request	8	8	6	3	25			
5. Rough Rice, Long Grain	Office	<u>8-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-20</u>	<u>20-26</u>	<u>All</u>		
	Crowley	5	8	9	9	4	35		
	Stuttgart	7	10	10	10	8	45		
6. Rough Rice, Medium Grain	Office	<u>8-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-21</u>	<u>21-28</u>	<u>All</u>		
	California	4	6	6	5	4	25		
	Crowley	2	2	2	2	2	10		
	Stuttgart	4	6	6	6	3	25		
	DIOO	2	6	6	4	2	20		
7. Sorghum	Office	<u>8-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-18</u>	<u>18-24</u>	<u>All</u>		
	DIOO	8	14	16	15	7	60		
8. Soybeans	Office	<u>5-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-18</u>	<u>18-24</u>	<u>All</u>	
	Toledo	5	6	7	6	4	2	30	
	DIOO	16	18	18	18	12	8	90	
9. Sunflower Seed, Oil Type	Office	<u>4-8</u>	<u>8-10</u>	<u>10-12</u>	<u>12-16</u>	<u>16-20</u>	<u>20-28</u>	<u>All</u>	
	DIOO	6	15	18	18	8	5	70	
10. Wheat, Durum	Office	<u>4-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-21</u>	<u>21-30</u>	<u>All</u>	
	California	1	3	3	1	1	1	10	
	DIOO	8	14	13	10	4	1	50	

Table 1. Sample collection assignments, 2019 Crop Year

		Moisture Range (%)					All
		<u>6-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-21</u>	
11. Wheat, Hard Red Spring	Office						
	Washington	2	3	2	2	1	10
	DIOO	6	15	17	15	7	60
		Moisture Range (%)					All
		<u>6-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-22</u>	
12. Wheat, Hard Red Winter	Office						
	Washington	2	2	2	2	2	10
	DIOO	4	16	17	16	7	60
		Moisture Range (%)					All
		<u>6-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>		
13. Wheat, Hard White	Office						
	California	1	2	1	1	5	
	Washington	3	5	5	2	15	
	DIOO	5	11	11	3	30	
		Moisture Range (%)					All
		<u>7-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-24</u>	
14. Wheat, Soft Red	Office						
	Toledo	4	4	4	4	4	20
	DIOO	8	11	12	11	8	50
		Moisture Range (%)					All
		<u>7-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-24</u>	
15. Wheat, Soft White	Office						
	Washington	4	7	8	7	4	30
	DIOO	4	7	8	7	4	30
		Moisture Range (%)					All
		<u>6-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-17</u>	<u>17-20</u>	
16. Lima Beans, Baby	Office						
	California	6	7	4	2	1	20
		Moisture Range (%)					All
		<u>8-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-17</u>	<u>17-20</u>	
17. Beans, Garbanzo	Office						
	Moscow	3	3	2	2	0	10
	DIOO	4	4	4	2	1	15
		Moisture Range (%)					All
		<u>6-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-17</u>	<u>17-20</u>	
18. Beans, Small Red	Office						
	Washington	2	4	3	1	0	10
	DIOO	2	3	3	1	1	10
		Moisture Range (%)					All
		<u>6-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-17</u>	<u>17-20</u>	
19. Peas, Mottled	Office						
	Moscow	5	4	4	1	1	15
	DIOO	2	2	1	0	0	5

Table 1. Sample collection assignments, 2019 Crop Year

		Moisture Range (%)					
		<u>6-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-17</u>	<u>17-20</u>	<u>All</u>
20. Peas, Smooth Dry	Office						
	Washington	1	2	2	0	0	5
	Moscow	3	3	3	1	0	10
	DIOO	1	1	1	1	1	5
		Moisture Range (%)					
		<u>6-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-17</u>	<u>17-20</u>	<u>All</u>
21. Peas, Wrinkled Dry	Office						
	Washington	3	4	1	1	1	10
	Moscow	3	4	3	0	0	10
		Moisture Range (%)					
		<u>8-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-20</u>	<u>20-26</u>	<u>All</u>
22. Rice, Long Grain Milled	Office						
	Crowley	6	6	5	3	0	20
	Stuttgart	8	8	7	5	2	30
		Moisture Range (%)					
		<u>8-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-20</u>	<u>20-26</u>	<u>All</u>
23. Rice, Medium Grain Milled	Office						
	Stuttgart	3	4	5	2	1	15
	Crowley	3	3	2	1	1	10
	California	2	2	4	1	1	10
	DIOO	1	1	1	1	1	5
		Moisture Range (%)					
		<u>8-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-20</u>	<u>20-26</u>	<u>All</u>
24. Rice, Long Grain Brown Rice	Office						
	Crowley	2	3	3	2	0	10
	(Parboiled) Stuttgart	5	5	5	3	2	20
		Moisture Range (%)					
		<u>8-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-20</u>	<u>20-26</u>	<u>All</u>
25. Rice, Long Grain Milled (Parboiled)	Office						
	Crowley	5	4	3	2	1	15
	Stuttgart	5	4	3	2	1	15
		Moisture Range (%)					
		<u>8-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-20</u>	<u>All</u>
26. Rice, Medium Grain Milled (Parboiled)	Office						
	Crowley	3	3	2	1	1	10
	Stuttgart	3	3	2	1	1	10
	California	4	4	5	1	1	15
		Moisture Range (%)					
		<u>8-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-20</u>	<u>All</u>
27. Rice, Brewers Milled (Parboiled)	Office						
	Crowley	2	2	1	0	0	5
	Stuttgart	4	5	4	1	1	15
		Moisture Range (%)					
		<u>8-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-20</u>	<u>All</u>
28. Rice, Second Head Milled (Parboiled)	Office						
	Stuttgart	6	6	5	2	1	20

Table 1. Sample collection assignments, 2019 Crop Year

		Moisture Range (%)					<u>All</u>
		<u>6-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-18</u>	<u>18-22</u>	
29. Popcorn	Office						
	DIOO	1	1	1	1	1	5
	Special Request	3	4	4	2	1	15
		Moisture Range (%)					<u>All</u>
		<u>8-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-18</u>	<u>18-22</u>	
30. Bean, Yelloweye	Office						
	Special Request	6	6	4	2	2	20
		Moisture Range (%)					<u>All</u>
		<u>8-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>		
31. Oats, Hulless	Office						
	Special Request	7	7	4	2	20	
		Moisture Range (%)					<u>All</u>
		<u>7-10</u>	<u>10-12</u>	<u>12-14</u>	<u>14-16</u>	<u>16-21</u>	
32. Barley, Hulless	Office						
	Washington	4	4	2	0	0	10
	DIOO	4	4	2	0	0	10